

Hb13 Series Brain-NIRS System

Specifications

| | |
|------------------------|---|
| Model | Hb133 |
| Measurement technology | NIRS (Near-infrared Spectroscopy), SRS (Spatially-Resolved Spectroscopy) |
| Light source | 2 colors of LEDs Average power: 1 mW or less Peak wavelength: 770±10 nm, 830±10 nm |
| Light detectors | Photodiodes |
| Sensor board | FPC (Flexible Printed Circuits) NOTE: Inappropriate handling may damage the circuits. |
| Number of channels | 2 channels (fixed) |
| Output data format | CSV format |
| Sampling rate | 0.1, 0.2, 0.5, 1.0 (sec) |
| Measurement items | Regional Oxygen Saturation (rSO ₂ : %) Oxyhemoglobin (Oxy-Hb: mM) Deoxyhemoglobin (Deoxy-Hb: mM) Triaxial acceleration (X, Y, Z, max. 2G) |
| Power source | 2 AAA batteries, 3V The faster the sampling rate is, the faster the battery will be eaten up. |
| Wireless connection | Bluetooth-LE (MBN52832) Within 5 m Foreign certificates: CE (Europe), FCC (US), IC (Canada), KC (Korea) |
| Measurement program | Windows Hb133.exe |
| Supported OS | Windows 8.1, 10 |
| Accessories | Bandana (2 sizes: L/M and M/S), SD card |
| Weight | Approx. 100 g (including batteries) |
| Included software | NIRS Data Analysis program |



| Product name | Model | Price(JPY/unit) (tax included) |
|-----------------------------|-------|-----------------------------------|
| Brain Activity Monitor* | Hb133 | 198,000 |
| Bandana (2 sizes: L/M, M/S) | Hb-b | 4,400 |

*Includes accessories

- In a bright environment, it may not be possible to measure correctly due to the influence of external light.
- Attach the sensor firmly to the skin. Hair between the sensor and skin interferes the measurement.
- Bandanas are available in two sizes.
- The program and other information will be provided with SD card.



astem

NOTES

- This device cannot be used for medical treatment. If you wish to use it, please obtain approval from the ethics committee of the university or hospital.
- You can use this product to develop and sell the application. We will provide the necessary information free of charge, so please feel free to contact us.
- The light measurement and correction algorithms for this product are based on technology owned by Shizuoka University and implemented by us under an exclusive license.

Manufacturer/Distributor

Controlled medical devices marketing license holder
Highly controlled medical devices marketing license holder
rental business

Astem Co.,Ltd.
Shimaya bldg.3F,2-14-6 Mizonokuchi,Takatsu-ku,
Kawasaki,Kanagawa 213-0001,Japan
TEL +81-44-833-8453 FAX +81-44-833-8456
Web: <https://astem-jp.com/>
E-mail: info@astem-jp.com (sales dept.)